

.

From: cag_support-na@agilent.com
Sent: Tuesday, October 09, 2012 10:55 AM
To: Hanchett, James (DPH)
Subject: FW: G1530A error / Massachusetts State Police / James Hanchett / support na jb

Importance: High

Hello,

So we can properly process your request, please forward us the following information –

- Your full name
- Your full company's name (no initials)
- Complete mailing address (no PO Boxes please)
- Your telephone number

After we receive your information, we will assign your request to a representative to contact you about your request.

Thank you.

*Helen Thomas
Agilent Technologies Inc
Customer Care Tech Coordinator
8002279770*

From: chem@agilent.com [mailto:chem@agilent.com]
Sent: Tuesday, October 09, 2012 9:33 AM
To: SALES-NA,CAG (A-USA,exgen1)
Subject: G1530A error / Massachusetts State Police / James Hanchett / support na jb

Name: James Hanchett Jim

Email: james.hanchett@state.ma.us

Country: United States

Telephone Number: 413-545-2607

Company Name: Massachusetts State Police

Product Group: Mass Spectrometry

Product Model Number: G1530A

Serial Number: US00022936

Software Revision: G1701EA E.02.00

Question Related To: Software

Comments: When running an auto sequence, fatal sequence error is detected and aborts. Error message 73 occurs and screen prints File 51 not open for OUTPUT or APPEND.

[Click here to mark this enquiry as closed](#)

message sent: 10/9/2012 7:33:24 AM

IP address: 192.25.91.52

Referrer: http://www.chem.agilent.com/en-US/Contact-us/_layouts/agilent/contactusquery.aspx?XsnLocation=/FormServerTemplates/ContactUsQueryRequest.xsn&Source=/en-US/Contact-Us/Pages/ContactUs.aspx&DefaultItemOpen=1&m=i
